

**FREE LIQUIDS
EPA 9095 B**

Page 1 of 1

Facility Name: _____ VELAP ID: _____

Assessor Name: _____ Analyst Name: _____ Inspection Date: _____

Relevant Aspect of Standards**Method
Reference****Y****N****N/A****Comments**

Records Examined: SOP Number/ Revision/ Date _____ Analyst: _____

Sample ID: _____ Date of Sample Preparation: _____ Date of Analysis: _____

Are conical paint filters with mesh number 60 +/- 5 %
(fine meshed size) used? (Available at local paint stores.)

4.1

Is 100 gram or mL representative sample used?

6.0

Note: materials such as sorbent pads or pillows should
be cut into small pieces (< 1 cm).

7.3

Note: Brittle materials > 1 cm may be lightly crushed or
cut.

7.3

Was sample allowed to drain for 5 minutes?

7.5

If any portion of the material collects in the graduated
cylinder was the material deemed to contain free liquids?

7.6

Notes/ Comments: